



Departamenti i Prokurimit

Prishtinë: datë, 03.01.2020

Zyrtari Përgjegjës i Prokurimit i Komunës së Prishtinës, në mbështetje të nenit 108/A (paragrafi 10.1) të Ligjit për Prokurimin Publik nr. 04/L-042, i ndryshuar dhe plotësuar me ligjin Nr. 04/L-237, ligjin Nr. 05/L-068 dhe ligjin Nr. 05/L-092, duke u bazuar në nenin 8 (paragrafi 8.1.2) të Rregullave për Parashtrimin e Ankesave të aprovuara nga Komisioni Rregullativ i Prokurimit Publik, duke vendosur sipas kërkesës për ri-shqyrtim të datës 21.12.2019 të parashtruar nga Operatori Ekonomik "INTERLAB SHPK për aktivitetin: "Furnizim me paisje mjekesore për QKMF dhe AMU-ritender-pjesa I" dhe me numër të prokurimit: **616 -19-8612-111** nxjerr:

V E N D I M

1. Refuzohet kërkesa për ri-shqyrtim e Operatori Ekonomik "INTERLAB SHPK" e datës 31.12.2019 lidhur me aktivitetin e prokurimit "Sherbimet e veçanta kontraktuale për perfaqësim ligjor për nevojat e Komunes së Prishtinës" dhe me numër të prokurimit: **616 -19-9412-221**, kundër Njoftimit për Vendimin e AK-se për dhenje të kontratës-pjesa e dyte të AK-së, të datës 26.12.2019, të publikuar në web-faqe të KRPP-së.

1. Mbetet në fuqi Vendimi për dhenje të kontratës për pjesën e dytë publikuar me datë: 26.12.2019;
2. Ky vendim hyn në fuqi ditën e nënshkrimit dhe do të publikohet për të gjitha palët në platformën elektronike në adresën : <https://e-prokurimi.rks-gov.net>.

A r s y e t i m

Operatori Ekonomik "Lego SHPK" me datë 31.12.2019 ka parashtruar kërkesën për ri-shqyrtim, lidhur me aktivitetin e prokurimit me titull: "Furnizim me paisje mjekesore për QKMF dhe AMU-ritender-pjesa I" dhe me numër të prokurimit: **616 -19-8612-111**.

Autoriteti Kontraktues, me datë 03.01.2020, ka shqyrtuar kërkesën në fjalë dhe ka konstatuar si vijon:

- Pretendimet janë të pa bazuara pasi që Autoriteti Kontraktues ka analizuar me kujdes tenderin e pretenduesit dhe ka konstatuar se rekomandimet e komisionit të vlerësimit janë të bazuara për arsyet në vijim:
- Komisioni me të drejte ka konstatuar se pretenduesi nuk ka ofruar deshmi të mjaftueshme për furnizime të ngjajshme në vlerën e kërkuar, pasi që deshmitë e e

dorezuara kane te bejne kryesisht me materiale shpenzuese,kemikalie,qelqurina,paisje laboratorike,paisje per veterinari dhe paisje te TI-se,ndersa lidhur me kerkesat teknike nuk ka asnje koment nga komisioni;

-Lidhur me pretendimet se tenderuesi i propozuar per kontrate nuk ka deshmi per pozicionet 1,2 dhe 6 poashtu jane te pa bazuara pasi qe ne tenderin e tij eshte paraqitur katalogu per artikujt,se bashku me specifikat dhe deklaratat e prodhuesit per konformitetin e tyre me specifikimet teknike te kerkuara.

Ne baze te sqarimeve paraprake u vendos si ne Dispozitiv te Vendimit.

Zyrtarja e prokurimit qe administron kete lende duhet te publikoj kete Vendim ne E Prokurim.

Këshilla juridike:

Kundër këtij vendimi, ankuesi mund të paraqesë ankesë pranë OSHP-së. **Ankesa prane OSHP-se duhet te dorezohet brenda 10 (dhjetë) diteve pas pranimi te ketij vendimi.**

Shtese:Deshmite nga katalogu i tenderuesit "Made Kos".


Zyrtari përgjegjës i Prokurimit



EC Declaration of Conformity



Manufacturer:

ARI Medical Technology Co., Ltd.

Add: Block A, Lotus Industrial Park, Economic Development Zone, Hefei City, Anhui, China.

According to EC Declaration of Conformity set out in Annex VII of the Council Directive 93/42/EEC (MDD93/42/EEC) concerning medical devices for the following products Series:

<u>Item No</u>	<u>Products Name</u>	<u>UMDNS Code</u>	<u>Classification</u>
DEF-9000A/B/C/D	Defibrillator	11129	Class III
DEF-8000A/B/C/D			
AED-7000/7000+			
DEF-7000			

And declare that:

- The product meets the essential requirements set out in MDD93/42/EEC Annex VII


The manufacturing site for the above products comply with EN ISO13485:2016 Quality Management System Requirements for Medical Devices.

ARI Medical Technology Co., Ltd.

Roy Wang

Sales Director

For and on behalf of
ARI Medical Technology Co., Limited
合肥奥立瑞医疗科技有限公司


.....
Authorized Signature(s)

Date: July 31, 2019



Product Service

Certificate

No. Q5 016389 0025 Rev. 00

Holder of Certificate: **VBM Medizintechnik GmbH**

Einsteinstr. 1
72172 Sulz a. N.
GERMANY

Facility(ies):

VBM Medizintechnik GmbH
Einsteinstr. 1, 72172 Sulz a. N., GERMANY

Certification Mark:



Scope of Certificate:

Design and development, production and distribution of medical products for Airway Management, Emergency, Accessories for Anesthesia and Intensive Care as well as Tourniquets

Applied Standard(s):

EN ISO 13485:2016
Medical devices - Quality management systems - Requirements for regulatory purposes (ISO 13485:2016)
DIN EN ISO 13485:2016

The Certification Body of TÜV SÜD Product Service GmbH certifies that the company mentioned above has established and is maintaining a quality management system, which meets the requirements of the listed standard(s). See also notes overleaf.

Report No.:

713117338

Valid from:

2018-11-19

Valid until:


2021-11-18

Date,

2018-11-19

Stefan Preiß

AM-12 Patient Monitor (12.1 inch)



Feature

1. 12.1-inch Six Standard parameters: ECG, RESP, NIBP, SPO2, 2_TEMP, PR/HR
2. 13 Type Arrhythmic Analysis, Multi_Lead ECG Waveforms Display in Phase, Real time S_T segment analysis, pacemaker detection Drug calculation and titratitable;
3. Efficient resistance to interference of defibrillator and electrosurgical cautery;
4. SPO2 can testing for 0.1% Weak;
5. RA-LL impedance Respiration;
6. Trend Coexist Display;
7. OxyCRG DyNamic View Display;
8. Bed to Bed view Display;
9. Networking capacity and nurse calling system;
10. Options: Central Monitoring System, IBP, EtCO2, Recorder, 12-Channel ECG, Touch Screen, VGA;
11. Built-in rechargeable battery 4400mAh;
12. 12.1" high resolution color TFT LCD display;
13. Large volume of tabular and graphic trends information storage and easy to recall;
14. Anti-ESU, anti- defibrillator;
15. Capture dynamic waveforms.

Technical Specification

ECG

Lead Mode: 5 Leads (R, L, F, N, C or RA, LA, LL, RL, V)

Lead selection: I, II, III, avR, avL, avF, V,
Waveform: 2 ch
Lead mode: 3 Leads (R, L, F or RA, LA, LL)
Lead selection: I, II, III,
Waveform: 1 ch
Gain: '2.5mm/mV, '5.0mm/mV, '10mm/mV, '20mm/mV, auto

Heart Rate and Alarm

Range

Adult: 15 ~ 300 bpm

Neo/Ped: 15 ~ 350 bpm

Accuracy: $\pm 1\%$ or ± 1 bpm, which great

Resolution: 1 bpm

Sensitivity: > 200 (μ V P-P)

Differential Input Impedance: > 5 M Ω

CMRR

-Monitor: > 105 dB

-Operation: > 105 dB

-Diagnosis: > 85 dB

Electrode offset potential: ± 300 mV

Leakage Current: < 10 μ A

Baseline Recovery: < 3 S after Defi.

ECG Signal Range: ± 8 mV (V_{p-p})

Bandwidth

-Surgery: 1 ~ 15 Hz

-Monitor: 0.5 ~ 35 Hz

-Diagnostic: 0.05 ~ 100 Hz

Calibration Signal: 1 (mV p-p), Accuracy: $\pm 5\%$

ST Segment Monitoring Range: Measure and Alarm -2.0 ~ +2.0 mV

ARR Detecting

Type: ASYSTOLE, VFIB/VTAC, COUPLET, BIGEMINY, TRIGEMINY, R ON T, VT >2 , PVC, TACHY, BRADY, MISSED BEATS, PNP, PNC

Alarm: Available

Review: Available

Respiration

Method: Impedance between R-F (RA-LL)

Differential Input Impedance: > 2.5 M Ω

Measuring Impedance Range: 0.3~5.0 Ω

Base line Impedance Range: 0 – 2.5 K Ω

Bandwidth: 0.3 ~ 2.5 Hz

Resp. Rate

Measuring and Alarm Range

-Adult: 0 ~ 120 rpm

-Neo/Ped: 0 ~ 150 rpm
Resolution: 1 rpm
Accuracy: ± 2 rpm
Apnea Alarm: 10 ~ 40 S

NIBP

Method: Oscillometric
Mode: Manual, Auto, STAT
Measuring Interval in AUTO Mode: 1, 2, 3, 4, 5, 10, 15, 30, 60, 90, 120, 180, 240, 480 (Min)
Measuring Period in STAT Mode: 5 Min
Pulse Rate Range: 40 ~ 240 bpm
Alarm Type: SYS, DIA, MEAN
Measuring and alarm range
Adult Mode
-SYS: 40 ~ 270 mmHg
-DIA: 10 ~ 215 mmHg
-MEAN: 20 ~ 235 mmHg
Pediatric Mode
-SYS: 40 ~ 200 mmHg
-DIA: 10 ~ 150 mmHg
-MEAN: 20 ~ 165 mmHg
Neonatal Mode
-SYS: 40 ~ 135 mmHg
-DIA: 10 ~ 100 mmHg
-MEAN: 20 ~ 110 mmHg
Resolution Pressure: 1mmHg
Accuracy Pressure Maximum Mean error: ± 5 mmHg
Maximum Standard deviation: ± 8 mmHg
Overpressure Protection
-Adult Mode: 297 ± 3 mmHg
-Pediatric Mode: 240 ± 3 mmHg
-Neonatal Mode: 147 ± 3 mmHg

SpO2

Measuring Range: 0 ~ 100 %
Alarm Range: 0 ~ 100 %
Resolution: 1 %
Accuracy: 70% ~ 100% ± 2 %; 0% ~ 69% unspecified
Actualization interval: about 1 Sec.
Alarm Delay: 10 Sec.
Pulse Rate
Measuring and Alarm Range: 20~300bpm
Resolution: 1bpm

Accuracy: ± 2 bpm

Temperature

Channel: 2

Measuring and Alarm Range: 0 ~ 50 °C

Resolution: 0.1°C

Accuracy: ± 0.1 °C

Actualization interval: about 1 Sec.

Average Time Constant: <10 Sec.

IBP(optional)

Label: ART, PA, CVP, RAP, LAP, ICP, P1, P2

Measuring and alarm range

-ART: 0 ~ 300 mmHg

-PA: -6 ~ 120 mmHg

-CVP/RAP/LAP/ICP: -10 ~ 40 mmHg

-P1/P2: -10 ~ 300 mmHg

Press Sensor

Sensitivity: 5 uV/mmHg

Impedance: 300-3000Ω

Resolution: 1 mmHg

Accuracy: $\pm 2\%$ or ± 1 mmHg, which great

Actualization interval: about 1 Sec

Standard Packing List

5-lead ECG cable (including RESP)

Adult finger SpO2 sensor

Adult NIBP cuff

NIBP extension tube

TEMP sensor

AC power adaptor

ECG electrodes

Operation instruction

Optional Function

Central Monitoring System, IBP, EtCO2, Recorder, 12-Channel ECG, Touch Screen, VGA



Add: Block B, CC Park, No.728 Lanzhou Road, Baohe Industrial Zone, Hefei, China.

Tel: 0086-551-63802963

Fax: 0086-551-63802920

Web: www.ari-cn.com

Email: sales@ari-cn.com



EC Declaration of Conformity



Manufacturer:

ARI Medical Technology Co., Ltd

Block A, Lotus Industrial Park, Economic Development Zone, Hefei City, Anhui, China

According to EC Declaration of Conformity set out in Annex VII of the Council Directive 93/42/EEC (MDD93/42/EEC) concerning medical devices for the following products Series:

<u>Item No</u>	<u>Products Name</u>	<u>UMDNS Code</u>	<u>Classification</u>
Space Blankets			
Spine board			
Scoop stretcher			
Cervical collars			
Self-loader	First-Aid Apparatus	16630	Class I
Stretchers			
Traction Splints			
First-aid Bag			
Resuscitator			


And declare that:

- The product meets the essential requirements set out in MDD93/42/EEC Annex VII

The manufacturing site for the above products comply with EN ISO13485:2016 Quality Management System Requirements for Medical Devices.

ARI Medical Technology Co., Ltd
Roy Wang
Sales Director

For and on behalf of
ARI Medical Technology Co., Limited
合肥奥立瑞医疗科技有限公司


.....
Authorized Signature(s)

Date: August 29, 2019



Benannt durch/Designated by
Zentralstelle der Länder
für Gesundheitsschutz
bei Arzneimitteln und
Medizinprodukten
www.zlg.de
ZLG-BS-244.10.08



Product Service

EC Certificate

Full Quality Assurance System

Directive 93/42/EEC on Medical Devices (MDD), Annex II excluding (4)
(Devices in Class IIa, IIb or III)

No. G1 016389 0017 Rev. 01

Manufacturer:

VBM Medizintechnik GmbH

Einsteinstr. 1
72172 Sulz a. N.
GERMANY

Facility(ies):

VBM Medizintechnik GmbH
Einsteinstr. 1, 72172 Sulz a. N., GERMANY

Product Category(ies): Tourniquet Systems and Pressure Infusors with
cuffs, Medical Devices for Trans Tracheal
Ventilations, Laryngeal Tubes, Respiration Sets,
Respiration accessories (except class I),
Rectal Tampon by Roche (RTR)

The Certification Body of TÜV SÜD Product Service GmbH declares that the aforementioned manufacturer has implemented a quality assurance system for design, manufacture and final inspection of the respective devices / device categories in accordance with MDD Annex II. This quality assurance system conforms to the requirements of this Directive and is subject to periodical surveillance. For marketing of class III devices an additional Annex II (4) certificate is mandatory. See also notes overleaf.

Report No.:

713137731_1

Valid from:

2018-10-21

Valid until:

2023-10-20

Date,

2018-10-18



Stefan Preiß

Lot item 8

Laryngeal Tube

Intubating Laryngeal Tube iLTS-D

VBM's 3rd generation supraglottic airway device



Up to 20% of all emergency intubations are estimated to be difficult to manage.¹⁾ Having the right system available is the cornerstone of effective airway management in an emergency. Airway algorithms recommend the usage of supraglottic airway devices (SAD) as options to overcome difficult scenarios.

In order to prevent aspiration, the NAP4 report prescribes the use of devices with gastric access.

SAD with intubating capability provide a way to reach a definitive airway (ET Tube placement).

As VBM's 3rd generation SAD the iLTS-D compiles all essential features required to achieve and protect an airway:

Ventilation – wide airway section to optimise gas flow and low pressure cuffs to maximise sealing performance (< 60 cmH₂O)

Drain Tube – allows the insertion of a gastric tube or suction catheter

Intubation – special design of ventilation lumen to enable fiberoptic insertion of an ET Tube

¹⁾ Walls, Ron MD et al. "Manual of Emergency Airway Management" (2012): 4th edition, chapter 2, p. 9.

Features



Efficient sizing


The iLTS-D is the ideal solution to save space in emergency bags and crash carts.

Only two sizes match all patients ≥ 125 cm.

2.5/3: 125 – 155 cm

4/5: ≥ 155 cm

Laryngeal Tube



Ventilation

The design of the ventilation section optimises gas flow and prevents airway obstruction from a downfolded epiglottis.

Thin walled cuffs guarantee a maximum airway leak pressure at low cuff pressure (< 60 cmH₂O).




Drain Tube

The drain tube can accommodate a large gastric tube or suction catheter:

2.5/3: max. 16 Fr

4/5: max. 18 Fr



Intubation

The iLTS-D secures the airway in emergency and enables fiberoptic placement of an ET Tube without compromising patency of supraglottic ventilation.

2.5/3: ET Tube max. 6.5 mm I.D.

4/5: ET Tube max. 8.0 mm I.D.

Order Information

Intubating Laryngeal Tube iLTS-D

For single use, sterile

Size	Patient	Single Set 1x iLTS-D 1x Syringe	Set of 10 10x iLTS-D	Intubation Set 1x iLTS-D 1x ET Tube with Stabilizer 1x Syringe
2.5/3	125–155 cm	REF 32-08-123-1	REF 32-08-023-1	REF 32-08-309-1
4/5	≥ 155 cm	REF 32-08-145-1	REF 32-08-045-1	REF 32-08-209-1

ET Tube with Stabilizer

Armored, for single use, sterile

Size	For iLTS-D Size	REF	Box
5.5 mm I.D.	2.5/3	31-40-055-1	10
7.5 mm I.D.	4/5	31-40-075-1	10



Application video: iLTS-D



Benannt durch/Designated by
Zentralstelle der Länder
für Gesundheitsschutz
bei Arzneimitteln und
Medizinprodukten
www.zlg.de
ZLG-BS-244.10.08



Product Service

EC Certificate

Full Quality Assurance System

Directive 93/42/EEC on Medical Devices (MDD), Annex II excluding (4)
(Devices in Class IIa, IIb or III)

No. G1 016389 0017 Rev. 01

Manufacturer:

VBM Medizintechnik GmbH

Einsteinstr. 1
72172 Sulz a. N.
GERMANY

Facility(ies):

VBM Medizintechnik GmbH
Einsteinstr. 1, 72172 Sulz a. N., GERMANY

Product Category(ies): Tourniquet Systems and Pressure Infusors with
cuffs, Medical Devices for Trans Tracheal
Ventilations, Laryngeal Tubes, Respiration Sets,
Respiration accessories (except class I),
Rectal Tampon by Roche (RTR)

The Certification Body of TÜV SÜD Product Service GmbH declares that the aforementioned manufacturer has implemented a quality assurance system for design, manufacture and final inspection of the respective devices / device categories in accordance with MDD Annex II. This quality assurance system conforms to the requirements of this Directive and is subject to periodical surveillance. For marketing of class III devices an additional Annex II (4) certificate is mandatory. See also notes overleaf.

Report No.:

713137731_1

Valid from:

2018-10-21

Valid until:

2023-10-20

Date,

2018-10-18

Stefan Preiß